

# VERIFICATION OF CALIBRATION REPORT

JUN 22 2010

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130492</u> ✓	
Name <u>Shayne Calt</u>		ID# <u>4559</u> ✓	Date <u>5-15-10</u> ✓
A	Agency <u>AST</u>	Phone # <u>733-2256</u>	
Instrument Location <u>Willow AST</u>			
Alco S/N <u>X301581</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>500</u>
B	Alco Test Value Average <u>.083</u> ✓		
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco <u>.083</u> ✓	
Signature <u>[Signature]</u>			
(OVER)			

*[Signature]*  
6/23/10


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of June, 2010.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130492

MAY 15, 2010

OPERATOR'S NAME:

CALT SHAYNE M

OPERATOR'S NUMBER: 4559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

V

O.L. #: V

DEPT/AGENCY: TAL0

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .082

ALCO S/N: X201581

### --- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.70 in

ALCO TARGET .081 09:04

BLANK TEST .000 09:05

INTERNAL STANDARD VERIFIED 09:05

ALCO TV 29.70 in .083 09:05

BLANK TEST .000 09:06

SUBJECT SAMPLE .000 09:07

BLANK TEST .000 09:07

ALCO TV 29.70 in .083 09:08

BLANK TEST .000 09:09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130492

MAY 15, 2010

TIME 09:01

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.70 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`ab cde fghijklmno  
pqrstuvwxyz{|}~